

内分泌复合非定值质控品赋值表 (批号: 202306151)

项目	仪器型号	方法学	单位	水平 1		水平 2		水平 3	
				靶值	范围	靶值	范围	靶值	范围
T3	Roche cobas e411	电化学发光法	nmol/L	1.48	1.04-1.92	2.06	1.44-2.67	5.26	3.68-6.83
	Snibe MAGLUMI X3	电化学发光法	ng/ml	0.31	0.21-0.40	0.71	0.5-0.93	2.58	1.81-3.35
	CL-8000i	电化学发光法	nmol/L	0.78	0.55-1.02	1.23	0.86-1.6	4.00	2.8-5.2
T4	Roche cobas e411	电化学发光法	nmol/L	25.94	18.16-33.73	80.70	56.49-104.91	131.50	92.05-170.95
	Snibe MAGLUMI X3	电化学发光法	µg/dL	2.30	1.61-2.99	7.56	5.29-9.82	11.27	7.89-14.65
	CL-8000i	电化学发光法	nmol/L	18.49	12.94-24.03	66.35	46.45-86.26	112.46	78.72-146.2
FT3	Roche cobas e411	电化学发光法	pmol/L	5.24	3.66-6.81	7.25	5.07-9.42	23.59	16.51-30.67
	Snibe MAGLUMI X3	电化学发光法	pg/mL	2.24	1.57-2.92	3.22	2.25-4.18	12.20	8.54-15.86
	CL-8000i	电化学发光法	pmol/L	3.72	2.6-4.84	6.05	4.24-7.87	16.34	11.43-21.24
FT4	Roche cobas e411	电化学发光法	pmol/L	8.30	5.81-10.79	26.34	18.44-34.24	41.43	29-53.86
	Snibe MAGLUMI X3	电化学发光法	ng/dL	0.94	0.66-1.22	2.78	1.95-3.61	4.16	2.91-5.41
	CL-8000i	电化学发光法	pmol/L	10.11	7.08-13.14	31.22	21.85-40.58	44.62	31.23-58

TG	Roche cobas e411	电化学发光法	ng/mL	2.88	2.01-3.74	14.77	10.34-19.2	54.56	38.19-70.93
	Snibe MAGLUMI X3	电化学发光法	ng/mL	≤0.02	≤0.02	0.64	0.45-0.83	10.86	7.6-14.11
	CL-8000i	电化学发光法	ng/mL	0.20	0.14-0.25	1.77	1.24-2.3	9.83	6.88-12.78
TPOAb	Roche cobas e411	电化学发光法	IU/mL	7.93	5.55-10.31	28.18	19.73-36.63	50.71	35.5-65.92
	Snibe MAGLUMI X3	电化学发光法	IU/mL	15.37	10.76-19.98	28.80	20.16-37.44	35.08	24.56-45.61
	CL-8000i	电化学发光法	IU/ml	11.38	7.96-14.79	23.67	16.57-30.77	32.93	23.05-42.81
TGAb	Roche cobas e411	电化学发光法	IU/mL	29.41	20.58-38.23	60.57	42.4-78.74	79.16	55.41-102.91
	Snibe MAGLUMI X3	电化学发光法	IU/mL	85.90	60.13-111.67	113.83	79.68-147.98	131.67	92.17-171.17
	CL-8000i	电化学发光法	IU/mL	8.82	6.18-11.47	14.79	10.35-19.23	15.83	11.08-20.58
TSH	Roche cobas e411	电化学发光法	μIU/mL	0.05	0.04-0.07	3.70	2.59-4.81	19.64	13.75-25.53
	Snibe MAGLUMI X3	电化学发光法	μIU/mL	0.37	0.26-0.49	2.55	1.78-3.31	5.79	4.05-7.52
	CL-8000i	电化学发光法	mIU/L	0.07	0.05-0.093	4.28	2.993-5.558	24.90	17.433-32.376
β-HCG	Roche cobas e411	电化学发光法	mIU/mL	3.31	2.31-4.3	23.98	16.78-31.17	498.92	349.24-648.59
	Snibe MAGLUMI X3	电化学发光法	mIU/mL	6.48	4.53-8.42	32.95	23.07-42.84	757.83	530.48-985.18
	CL-8000i	电化学发光法	mIU/mL	2.05	1.43-2.66	20.73	14.51-26.95	341.01	238.71-443.31
FSH	Roche cobas e411	电化学发光法	mIU/mL	4.06	2.84-5.28	16.17	11.32-21.02	35.93	25.15-46.7

	Snibe MAGLUMI X3	电化学发光法	mIU/mL	4.67	3.27-6.07	17.83	12.48-23.18	47.38	33.17-61.6
	CL-8000i	电化学发光法	mIU/mL	5.20	3.64-6.76	20.02	14.01-26.02	54.61	38.23-71
LH	Roche cobas e411	电化学发光法	mIU/mL	3.28	2.3-4.27	12.96	9.07-16.84	39.90	27.93-51.87
	Snibe MAGLUMI X3	电化学发光法	mIU/mL	4.11	2.88-5.35	20.97	14.68-27.26	53.22	37.25-69.18
	CL-8000i	电化学发光法	mIU/mL	3.24	2.27-4.22	13.39	9.37-17.41	45.98	32.18-59.77
PRL	Roche cobas e411	电化学发光法	μIU/mL	83.73	58.61-108.85	252.03	176.42-327.64	886.73	620.71-1152.75
	Snibe MAGLUMI X3	电化学发光法	ng/mL	3.08	2.16-4	10.73	7.51-13.95	33.35	23.35-43.36
	CL-8000i	电化学发光法	ng/mL	5.25	3.68-6.83	12.09	8.46-15.71	40.92	28.64-53.19
E2	Roche cobas e411	电化学发光法	pg/mL	52.09	36.46-67.71	211.93	148.35-275.51	535.18	374.63-695.74
	Snibe MAGLUMI X3	电化学发光法	pg/mL	94.47	66.13-122.81	302.00	211.4-392.6	823.50	576.45-1070.55
	CL-8000i	电化学发光法	pg/mL	57.01	39.91-74.12	246.99	172.89-321.09	645.57	451.9-839.23
PROG	Roche cobas e411	电化学发光法	ng/mL	1.10	0.77-1.43	11.38	7.97-14.8	26.86	18.8-34.92
	Snibe MAGLUMI X3	电化学发光法	ng/mL	0.58	0.41-0.75	8.13	5.69-10.57	17.87	12.51-23.23
	CL-8000i	电化学发光法	ng/ml	2.74	1.92-3.56	34.12	23.88-44.35	95.92	67.15-124.7
TEST	Roche cobas e411	电化学发光法	ng/mL	0.42	0.29-0.54	3.66	2.56-4.76	8.53	5.97-11.08
	CL-8000i	电化学发光法	nmol/L	<0.10	<0.10	2.86	2-3.72	9.33	6.53-12.13

DHEA-S	Roche cobas e411	电化学发光法	μg/dL	48.69	34.08-63.29	183.52	128.46-238.57	581.40	406.98-755.82
	Snibe MAGLUMI X3	电化学发光法	μg/dL	38.10	26.67-49.53	139.50	97.65-181.35	362.83	253.98-471.68
17a-OH P	Snibe MAGLUMI X3	电化学发光法	ng/ml	≤0.035	≤0.035	2.87	2.01-3.73	10.14	7.1-13.18
25-OH VD	Snibe MAGLUMI X3	电化学发光法	ng/ml	48.43	33.9-62.96	47.23	33.06-61.4	71.35	49.95-92.76
	CL-8000i	电化学发光法	ng/mL	27.05	18.94-35.17	40.43	28.3-52.56	54.33	38.03-70.62
PTH	Roche cobas e411	电化学发光法	pg/mL	14.00	9.8-18.2	159.38	111.57-207.2	651.98	456.39-847.58
	Snibe MAGLUMI X3	电化学发光法	pg/mL	16.50	11.55-21.45	191.67	134.17-249.17	725.33	507.73-942.93
	CL-8000i	电化学发光法	pg/ml	12.33	8.63-16.02	160.48	112.33-208.62	599.50	419.65-779.35
TP1NP	Roche cobas e411	电化学发光法	ng/mL	13.20	9.24-17.16	24.96	17.47-32.45	224.65	157.26-292.05
	Snibe MAGLUMI X3	电化学发光法	ng/mL	10.25	7.18-13.33	22.97	16.08-29.86	179.67	125.77-233.57
Insulin	Roche cobas e411	电化学发光法	μU/mL	17.39	12.17-22.6	44.54	31.17-57.9	161.73	113.21-210.25
	Snibe MAGLUMI X3	电化学发光法	μIU/mL	15.38	10.77-20	40.92	28.64-53.19	155.67	108.97-202.37
	CL-8000i	电化学发光法	μIU/mL	22.61	15.83-29.4	55.09	38.56-71.61	211.71	148.19-275.22
C-pep	Roche cobas e411	电化学发光法	ng/mL	1.02	0.71-1.33	4.38	3.07-5.7	7.72	5.4-10.03
	Snibe MAGLUMI X3	电化学发光法	ng/mL	0.85	0.6-1.11	4.16	2.91-5.41	7.65	5.36-9.95
	CL-8000i	电化学发光法	ng/ml	0.70	0.49-0.91	3.73	2.61-4.84	7.33	5.13-9.52

VB12	Roche cobas e411	电化学发光法	pg/mL	349.17	244.42-453.92	652.52	456.76-848.27	1071.25	749.88-1392.63
	Snibe MAGLUMI X3	电化学发光法	pg/mL	221.00	154.7-287.3	453.67	317.57-589.77	775.33	542.73-1007.93
	CL-8000i	电化学发光法	pg/mL	240.90	168.63-313.17	536.09	375.26-696.91	1036.37	725.46-1347.28
FOL	Roche cobas e411	电化学发光法	ng/mL	4.58	3.2-5.95	6.56	4.59-8.53	11.84	8.29-15.39
	Snibe MAGLUMI X3	电化学发光法	ng/mL	0.89	0.63-1.16	2.15	1.5-2.79	2.51	1.76-3.26
	CL-8000i	电化学发光法	nmol/L	14.35	10.04-18.65	17.07	11.95-22.18	33.18	23.23-43.14
ACTH	Roche cobas e411	电化学发光法	pg/mL	11.11	7.78-14.44	56.49	39.54-73.43	812.80	568.96-1056.64
	Snibe MAGLUMI X3	电化学发光法	pg/mL	4.54	3.18-5.9	46.57	32.6-60.54	804.83	563.38-1046.28
CORT	Roche cobas e411	电化学发光法	nmol/L	115.28	80.7-149.87	437.77	306.44-569.1	1116.33	781.43-1451.23
	Snibe MAGLUMI X3	电化学发光法	ng/ml	70.52	49.36-91.67	206.33	144.43-268.23	514.50	360.15-668.85
	CL-8000i	电化学发光法	nmol/L	111.99	78.39-145.59	406.68	284.68-528.69	1034.49	724.14-1344.83
ALD	Snibe MAGLUMI X3	电化学发光法	pg/mL	9.80	5.44-10.51	69.70	64.2-75.2	1379.00	1168.8-1589.2
IGF-1	Snibe MAGLUMI X3	电化学发光法	ng/mL	26.07	18.25-33.89	52.83	36.98-68.68	152.17	106.52-197.82
hGH	Roche cobas e411	电化学发光法	pg/mL	1480.17	1036.12-1924.22	4678.67	3275.07-6082.27	12565.33	8795.73-16334.93
	Snibe MAGLUMI X3	电化学发光法	ng/ml	1.57	1.1-2.05	4.90	3.43-6.37	13.97	9.78-18.16

注：以上靶值及范围具有批特异性，仅供本批质控品参考。因所用技术、仪器、试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据，根据良好的实验室规范要求，实验室须遵循相关技术规范确立自己的靶值和可接受范围。