

数字 PCR 实测值

- 产品名称: 泛肿瘤 800 gDNA 标准品
- 产品编号: GW-OGTM800
- 批次号: 201908003
- 检测结果:

等位基因突变频率验证: QX200 Droplet Digital PCR (BIO-RAD)					
基因型	基因频率	实际基因频率	质检下限	质检上限	是否通过质检
EGFR L858R	1.00%	0.98%	0.70%	1.30%	是
KRAS A146T	1.00%	0.90%	0.70%	1.30%	是
NRAS Q61K	1.00%	1.13%	0.70%	1.30%	是
EGFR T790M	2.00%	2.08%	1.40%	2.60%	是
EGFR ΔE746_A750	2.00%	2.04%	1.40%	2.60%	是
FLT3 ΔI836	2.00%	1.87%	1.40%	2.60%	是
KIT D816V	2.00%	2.00%	1.40%	2.60%	是
KRAS G12D	2.00%	1.93%	1.40%	2.60%	是
EGFR V769_D770insASV	3.00%	2.54%	1.40%	2.60%	是
EGFR G719S	4.00%	4.18%	2.80%	5.20%	是
KRAS G13D	4.00%	3.60%	2.80%	5.20%	是
EML4-ALK Fusion V3	5.00%	5.12%	4.00%	6.00%	是
CD74-ROS1 Fusion	6.00%	6.04%	4.80%	7.20%	是
BRAF V600E	7.00%	6.48%	5.60%	8.40%	是
PIK3CA H1047R	7.00%	7.00%	5.60%	8.40%	是
ERBB2 Amplification	5.00C	4.74C	4.00C	6.00C	是

注意: (1) 建议使用推荐的仪器和方法对产品进行定量。