

## 数字 PCR 实测值

- 产品名称: 肿瘤 SNV 野生型 ctDNA 标准品
- 产品编号: GW-OCTM005
- 批次号: 202106001
- 检测结果:

等位基因突变频率验证: QX200 Droplet Digital PCR (BIO-RAD)					
基因型	基因频率	实际基因频率	质检下限	质检上限	是否通过质检
AKT1 E17K	0.00%	0.00%	0.00%	0.10%	Pass
BRAF V600E	0.00%	0.00%	0.00%	0.10%	Pass
EGFR G719S	0.00%	0.00%	0.00%	0.10%	Pass
EGFR L858R	0.00%	0.00%	0.00%	0.10%	Pass
EGFR T790M	0.00%	0.00%	0.00%	0.10%	Pass
EGFR V769_D770insASV	0.00%	0.00%	0.00%	0.10%	Pass
EGFR ΔE746_A750	0.00%	0.00%	0.00%	0.10%	Pass
KRAS A146T	0.00%	0.00%	0.00%	0.10%	Pass
KRAS G12D	0.00%	0.00%	0.00%	0.10%	Pass
KRAS G13D	0.00%	0.03%	0.00%	0.10%	Pass
NRAS Q61K	0.00%	0.01%	0.00%	0.10%	Pass
PIK3CA E545K	0.00%	0.01%	0.00%	0.10%	Pass
PIK3CA H1047R	0.00%	0.00%	0.00%	0.1%	Pass
CD74-ROS1 Fusion	0.00%	0.00%	0.00%	0.1%	Pass
EML4-ALK Fusion	0.00%	0.00%	0.00%	0.10%	Pass
ERBB2 Amplification	2.00 copies	1.64 copies	1.20 copies	2.80 copies	Pass
MET Amplification	2.00 copies	1.61 copies	1.20 copies	2.80 copies	Pass

注意: (1) 建议使用推荐的仪器和方法对产品进行定量。