

病原体血清学复合阴性非定值质控品靶值表
(批号: 202407001)

仪器型号	项目	单位	范围
CL-6200i CL-8000i	HBsAg	IU/mL	< 0.08
	Anti-HBs	mIU/mL	< 10
	HBeAg	PEI U/mL	< 0.1
	Anti-HBe	COI	> 1
	Anti-HBc	COI	> 1
	Anti-TP	COI	< 1
	HIV	COI	< 1
	Anti-HCV	COI	< 1
iFlash 3000	HBsAg	IU/mL	< 0.05
	Anti-HBs	mIU/mL	< 10
	HBeAg	COI	< 1
	Anti-HBe	COI	> 1
	Anti-HBc	COI	> 1
	Anti-TP	COI	< 1
	HIV	COI	< 1
	Anti-HCV	COI	< 1
maccura i3000	HBsAg	IU/mL	< 0.05
	Anti-HBs	mIU/mL	< 10
	HBeAg	IU/mL	< 0.1
	Anti-HBe	IU/mL	< 0.2
	Anti-HBc	IU/mL	< 0.6
	Anti-TP	S/CO	< 1
	HIV	S/CO	< 1
	Anti-HCV	S/CO	< 1
Cobas e601	HBsAg	COI	< 0.9
	Anti-HBs	IU/L	< 10
	HBeAg	COI	< 1
	Anti-HBe	COI	> 1
	Anti-HBc	COI	> 1
	Anti-TP	COI	< 1
	HIV	COI	< 1
	Anti-HCV	COI	< 1
AutoLumo A2000	HBsAg	IU/mL	< 0.05
	Anti-HBs	mIU/mL	< 10
	HBeAg	IU/mL	< 0.1
	Anti-HBe	PEI U/mL	< 0.15
	Anti-HBc	PEI U/mL	< 0.7
	Anti-TP	S/CO	< 1
	HIV	S/CO	< 1
	Anti-HCV	S/CO	< 1
ARCHITECT i4000SR	HBsAg	IU/mL	< 0.05
	Anti-HBs	mIU/mL	< 10
	HBeAg	S/CO	< 1

	Anti-HBe	S/CO	> 1
	Anti-HBc	S/CO	< 1
	Anti-TP	S/CO	< 1
	HIV	S/CO	< 1
	Anti-HCV	S/CO	< 1
MAGLUMI X3	HBsAg	IU/mL	< 0.05
	Anti-HBs	mIU/mL	< 10
	HBeAg	IU/mL	< 0.1
	Anti-HBe	IU/mL	< 0.18
	Anti-HBc	IU/mL	< 0.35
	Anti-TP	mIU/mL	< 1
	HIV	AU/mL	< 1
	Anti-HCV	AU/mL	< 5
Caris 200	HBsAg	IU/mL	< 0.05
	Anti-HBs	mIU/mL	< 10
	HBeAg	PEIU/mL	< 0.1
	Anti-HBe	PEIU/mL	< 0.15
	Anti-HBc	IU/mL	< 0.35
	Anti-TP	S/CO	< 1
	HIV	S/CO	< 1
	Anti-HCV	S/CO	< 1

注：以上靶值仅供本批质控品参考，不作为定值依据。因所用技术、仪器、试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据，根据良好的实验室规范要求，实验室须遵循相关技术规范确立自己的靶值和可接受范围。