

病原体血清学复合阴性非定值质控品靶值表
(批号: 202409001)

仪器型号	项目	单位	范围
maccura i3000	HBsAg	IU/mL	<0.05
	Anti-HBs	mIU/mL	<10
	HBeAg	IU/mL	<0.1
	Anti-HBe	IU/mL	<0.2
	Anti-HBc	IU/mL	<0.6
	Anti-TP	S/CO	<1
	HIV	S/CO	<1
	Anti-HCV	S/CO	<1
Roche cobas e601	HBsAg	COI	<0.9
	Anti-HBs	IU/L	<10
	HBeAg	COI	<1
	Anti-HBe	COI	>1
	Anti-HBc	COI	>1
	Anti-TP	COI	<1
	HIV	COI	<1
	Anti-HCV	COI	<1
AutoLumo A6200	HBsAg	IU/mL	<0.05
	Anti-HBs	mIU/mL	<10
	HBeAg	IU/mL	<0.1
	Anti-HBe	PEI U/mL	<0.15
	Anti-HBc	PEI U/mL	<0.7
	Anti-TP	S/CO	<1
	HIV	S/CO	<1
	Anti-HCV	S/CO	<1
Abbott i2000SR	HBsAg	IU/mL	<0.05
	HBsAg	mIU/mL	<10
	Anti-HBs	S/CO	<1
	HBeAg	S/CO	>1
	Anti-HBe	S/CO	<1
	Anti-HBc	S/CO	<1
	Anti-TP	S/CO	<1
	HIV	S/CO	<1
	Anti-HCV	IU/mL	<0.05
Snibe MAGLUMI X3	HBsAg	IU/mL	<0.05
	Anti-HBs	mIU/mL	<10
	HBeAg	IU/mL	<0.1
	Anti-HBe	IU/mL	<0.18
	Anti-HBc	IU/mL	<0.35
	Anti-TP	mIU/mL	<1
	HIV	AU/mL	<1
	Anti-HCV	AU/mL	<5

WANTAI caris200	HBsAg	IU/mL	<0.05
	Anti-HBs	mIU/mL	<10
	HBeAg	PE IU/mL	<0.1
	Anti-HBe	PE IU/mL	<0.15
	Anti-HBc	IU/mL	<0.35
	Anti-TP	S/CO	<1
	HIV	S/CO	<1
	Anti-HCV	S/CO	<1
Mindray CL6200i Mindray CL8000i	HBsAg	IU/mL	<0.08
	Anti-HBs	mIU/mL	<10
	HBeAg	PEI U/mL	<0.1
	Anti-HBe	COI	>1
	Anti-HBc	COI	>1
	Anti-TP	COI	<1
	HIV	COI	<1
	Anti-HCV	COI	<1
YHLO iFlash 3000	HBsAg	IU/mL	<0.05
	Anti-HBs	mIU/mL	<10
	HBeAg	COI	<1
	Anti-HBe	COI	>1
	Anti-HBc	COI	>1
	Anti-TP	COI	<1
	HIV	COI	<1
	Anti-HCV	COI	<1

注：以上靶值仅供本批质控品参考，不作为定值依据。因所用技术、仪器、试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据，根据良好的实验室规范要求，实验室须遵循相关技术规范确立自己的靶值和可接受范围。